

SY014-2 Two Wire 4 to 20mA Output Loadcell Amplifier

Operating Information

Loadcells fitted with SY014-2 amplifiers are supplied set up as detailed below unless special instructions are stated on the order. A System Certificate is supplied giving the amplifier output at each loadcell calibration point.

Supply voltage	Recommended	24.0V	
	Maximum	36.0V	
	Minimum	20.0V	
	Current limit	<1.0A	
	Ripple	<10mV RMS	
Standard output	Zero load	4.0mA	
	Full load	20.0mA	
Bi-directional output	Tension full load	4.0mA	
	Zero load	12.0mA	
	Compression full load	20.0mA	
Maximum loop resistance	24V supply	600Ω	
	20V supply	400Ω	
Calibration configuration	24V supply	250 Ω loop resistance	
Span setting accuracy	Standard output	±0.3% of span	
	Bi-directional output	±0.5% of span	
Non-linearity - Typical		±0.05% of full range	
Operating temperature range		0 to 50°C	

Connections

Function	Loadcell Cable PVC Wire Colour	Loadcell Cable Raychem Wire Colour	MS Connector Pin Number	Binder 723, Lemo 1B & DIN Connector Pin Number
+Loop	Red	Red	A	1
-Loop	Blue	Black	В	2
Screen	Orange	Orange	No connection	No connection

Use good quality screened cable for the output with the screen only earthed at one point. This will usually be at the electronics powering the amplifier. The connector options are only available on certain loadcells, please check the loadcell data sheets for availability.

The body of the loadcell must be connected to a good quality earth to maintain the EMC performance of the product.

THE SUPPLY CONNECTIONS MUST NOT BE REVERSED.

C E This product complies with the requirements of the European EMC directive.

Document number MN0016 Issue 6 Date 11:01:06

NOVATECH MEASUREMENTS LTD

83 CASTLEHAM ROAD, ST LEONARDS ON SEA, EAST SUSSEX, TN38 9NT, ENGLAND Tel: 01424 852744 email: info@novatechloadcells.co.uk Fax: 01424 853002 www.novatechloadcells.co.uk